

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVK-P

S.N.:

959

Date:

JUL 31 82



a) Output signal amplitude:

$\geq +15V$

b) Pulse width:

3 NSEC TO 100 NSEC

c) Rise time:

$\leq 100 \mu\text{SEC}$

d) Fall time:

$\leq 2 \text{ NSEC}$

e) PRF:

SEE FIG 1+2

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0V$

160 mA.

10 VOLTS/DIV

1.0 NSEC/DIV

200 KHZ.

[Signature]