


PULSE GENERATOR
PERFORMANCE CHECK

Model: **PAK-P**
S.N.: **960**
Date: **1627 31 82**



10 VOLTS/DIV.
1.0 NSEC/DIV.
200 KHz

- a) Output signal amplitude:
> +15V
- b) Pulse width:
3 NSEC TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE F14 1+2
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0V
130 mA.