


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVL-TN-1

S.N.: 967

Date: Aug 30 82



- a) Output signal amplitude:  
-230V
- b) Pulse width:  
5 TO 100 NSWC.
- c) Rise time:  
≤ 2 NSWC.
- d) Fall time:  
≤ 3 NSWC
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50 - 60 Hz
- h) DC OFFSET:  
+30 V,  
6 mA (max)
- i) Prop. DELAY!  
≤ 100 NSWC.