


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-A-N*

S.N.: *968*

Date: *AUG 31 82*



- a) Output signal amplitude:
0 TO -50V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 500 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20. KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*32 VOLTS/DIV
5 NSEC/DIV
10 KHZ.*

*+15 ± 1.0 V
260 mA (MAX)*