


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-AN-1-A-W

S.N.: 970

Date: 1944 31 82



- a) Output signal amplitude:
0 TO 100 V.
- b) Pulse width: *
0 TO 400 NSEC
- c) Rise time:
 ≤ 1 NSEC
- d) Fall time:
 ≤ 2 NSEC.
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

5 NSEC/DIV.
100 VOLTS/DIV
4 KHz

+15 ± 1.0 V
400 mA (MAX)

* REQUIRES EXTERNAL
DELAY LINE FOR
100 TO 400 NSEC
RANGE.