


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-A-P-TTL*

S.N.: *972*

Date: *OCT 17 82*



- a) Output signal amplitude:  
*0 to +15V*
- b) Pulse width:  
*2 nsec to 100 nsec*
- c) Rise time:  
*≤ 100 nsec*
- d) Fall time:  
*≤ 2 nsec*
- e) PRF:  
*SET P165 1,2*
- f) Jitter, stability:  
*O.K.*
- g) Prime power:

*1.0 nsec/div.  
10 VOLTS/div.  
40 KHz*

*+15 ± 1.0 V  
250 mA (MAX)  
*[Signature]**