


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVI-21**
S.N.: **976**
Date: **JUNE 24 82**



1 NSEC/DIV
32 VOLTS/DIV
40 KHZ

- a) Output signal amplitude:
 $\geq +40V$
- b) Pulse width:
1 TO 100 NSEC
- c) Rise time:
 $\leq 300 PSEC$
- d) Fall time:
 $\leq 2 NSEC$
- e) PRF:
505 F/1 1+2
- f) Jitter, stability:
OK
- g) Prime power:
**+15V \pm 1.0V,
130 mA OR
+135V \pm 5V,
5 mA.**