


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-1-N*

S.N.: *977*

Date: *Aug 22 82*



a) Output signal amplitude:

$\approx -250V$

b) Pulse width:

$\approx 4 \text{ NSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ VOLT}$

250 mA (MAX)

100 VOLTS/DIV

2.0 NSEC/DIV

40 KHZ