


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AL0-1*
S.N.: *981*
Date: *13 SEPT 1982*



- a) Output signal amplitude:
+25 TO +100 AMPS
- b) Pulse width:
25 NSEC
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1 KITE
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 V
185 mA (max)

64 AMP/DIV
10 NSEC/DIV
1 KITE

+15 ± 1.0 V
185 mA (max)
[Signature]