


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1*  
S.N.: *983*  
Date: *SEPT 13 82*



- a) Output signal amplitude:  
*+ 25 TO + 100 AMPS*
- b) Pulse width:  
*≈ 25 NSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 V.*

*64 AMPS / DIV.  
10 NSEC / DIV.  
1 KHz.*

*+15 ± 1.0 V.  
140 mA (MAX)*