


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1*
S.N.: *986*
Date: *SEPT 13 1982*



- a) Output signal amplitude:
+25 TO 100 AMPS
- b) Pulse width:
≈ 25 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*64 AMPS / DIV
10 NSEC / DIV
1 KHz*

*+15 ± 1.0 V.
125 mA (max)
*[Signature]**