


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVP-AU-HV2-P*

S.N.: *987*

Date: *SEPT 27 82*



*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*400 KHZ*



*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*400 KHZ*

a) Output signal amplitude:

*0 TO +20V*

b) Pulse width:

*0 TO 2 NSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 200 PSEC*

e) PRF:

*0 TO 1 MHz*

f) Jitter, stability:

*OK.*

g) Prime power:

*+15 ± 1.0 VOLTS,*

*225 mA (MAX)*

*[Signature]*