


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV7-AV-2-P-05-P3

S.N.: 991

Date: NOV 18 82



a) Output signal amplitude:

0 TO +10V

b) Pulse width:

5 TO 50  $\mu$ SEC

c) Rise time:

$\leq$  100 PSEC

d) Fall time:

$\leq$  135 PSEC

e) PRF:

0 TO 50 KHZ

f) Jitter, stability:

OK

g) Prime power:

110/220 V

50-60 HZ

10 V / DIV  
2  $\mu$ SEC / DIV  
40 KHZ