


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-N*

S.N.: *999*

Date: *AUG 23 82*



a) Output signal amplitude:

*> -20V*

b) Pulse width:

*≤ 0.5 nsec*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 VOLT*

*150 mA*

*(MAX)*

*16 VOLTS/DIV*

*1.0 nsec/DIV*

*400 kHz*