


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-N-D-PS*

S.N.: *1000*

Date: *SEPT 27 82*



- a) Output signal amplitude:  
*0 TO -5V*
- b) Pulse width:  
*0 TO 6 NSER*
- c) Rise time:  
*≤ 100 PSEER*
- d) Fall time:  
*≤ 135 PSE*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*110/220V*

*50-60 Hz*

*5 VOLTS/DIV*

*1.0 NSER/DIV*

*4.0 MHz*