


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-ENP*

S.N.: *1021*

Date: *OCT 12 82*



5 VOLTS / DIV
1.0 nSEC / DIV
1.0 MHz

- a) Output signal amplitude:
0 TO +5 V
- b) Pulse width:
0 TO 6 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 V
300 mA MAX

