


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-EW-P*

S.N.: *1024*

Date: *Oct 12 82*



*5 VOLTS/DIV
1.0 NS/DIV
1.0 MHz*

a) Output signal amplitude:

0 TO +5V

b) Pulse width:

0 TO 6 NS

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 V

300 mA (MAX)

MY

