


PULSE GENERATOR
PERFORMANCE CHECK

Model: 1A20-1
S.N.: 1027
Date: SEPT 13 1982



- a) Output signal amplitude:
+25 TO +100 AMPS
- b) Pulse width:
≈ 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 V
130 mA (MAX)

64 AMPS/DIV
10 NSEC/DIV
1 KHz

[Handwritten signature]