


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1*  
S.N.: *1028*  
Date: *SEPT 13 1982*



- a) Output signal amplitude:  
*+ 25 TO + 100 AMP*
- b) Pulse width:  
*≈ 25 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*64 AMP/DIV.*  
*10 NSEC/DIV*  
*1 KHZ.*

*+ 15 ± 1.0 V.*  
*140 mA (MAX)*