

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


AVO-3B-1

S.N.:

1042

Date:

Aug 22 82



a) Output signal amplitude:

+10 to +30 mVPS

b) Pulse width:

≤ 1.5 nsec.

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V.

175 mA (MAX)

20 AMP/DIV.

2.0 NSEC/DIV.

20 KHZ.