


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-67A-PS-P

S.N.: 1043

Date: OCT 31 82



- a) Output signal amplitude:
0 TO +200V*
- b) Pulse width:
50 NSEC TO 1 ~~ms~~ msec.
- c) Rise time:
≤ 10 NSEC.
- d) Fall time:
≤ 10 NSEC.
- e) PRF:
Checked 0 to 10 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 HZ

50 VOLTS / DIV
200 NSEC / DIV
1 KHZ
RL = 50 Ω

* FOR LASER DIODE LOAD WE RECOMMEND:

