


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-N*
S.N.: *1046*
Date: *OCT 17 82*



- a) Output signal amplitude:
0 to 10V
- b) Pulse width:
0 to 4 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 135 psec
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

1.0 nsec/div
10 volts/div
400 kHz

+15 ± 1.0 V
200 mA (max)

[Signature]