


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-A-N*
S.N.: *1047*
Date: *OCT 17 82*



- a) Output signal amplitude:
0 TO -15
- b) Pulse width:
2.0 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE PWS 1+2
- f) Jitter, stability:
OK
- g) Prime power:

1.0 NSEC/DIV.
10 VOLTS/DIV.
40 KHz.

+15 ± 1.0 VOLT.
200 mA 