


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-AV-P-C.*

S.N.: *1049*

Date: *DEC 17 82*



- a) Output signal amplitude:  
*0 TO +15 VOLTS.*
- b) Pulse width:  
*0 TO 50 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 50 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power: *\**

*1 NSEC/DIV*  
*10 VOLTS/DIV*  
*40 KHZ.*

*120/240 V,*  
*50-60 HZ.*

*UNIT SHIPPED*  
*WITH VOLTAGE*  
*SELECTOR (AND*  
*IN 240 V*  
*POSITION.)*