


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P*

S.N.: *1053*

Date: *DEC 21 82*



- a) Output signal amplitude:
0 TO +15 V
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 23 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
4.0 MHz

+24 ± 1.0 V
450 mA (MAX)
[Signature]