



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P-D-PS*
S.N.: *1066*
Date: *DEC 20 82*



*10 VOLTS/DIV
1.0 NSEC/DIV
4.0 MHz*



*10 VOLTS/DIV
1.0 NSEC/DIV
25 MHz*

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*110/220V
50-60 Hz*