


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-1-C-P

S.N.: 1067

Date: OCT 25 82



a) Output signal amplitude:

0 TO +10 V

b) Pulse width:

≤ 135 PSEC

c) Jitter, stability:

OK

d) Prime power:

110 / 220 V

50-60 Hz

e) HIGH TEMP OPERATION.

OPERATION CHECKED
TO +130°F.

1.0 NSEC / DIV
10 VOLTS / DIV
400 kHz