


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AV-1-N-TTL-01-PS*

S.N.: *1069*

Date: *NOV 30 82*



- a) Output signal amplitude:
0 TO 10V
- b) Pulse width:
0 TO 4 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 nSEC/DIV
400 kHz.*

*120/240 V 50-60 Hz
THIS UNIT SHIPPED
WITH THE VOLTAGE
SELECTOR CHANGED IN
THE 240V POSITION*