


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S-1-N-A-PS

S.N.: 1073

Date: NOV 30 82



a) Output signal amplitude:

0 TO - 10 V

b) Pulse width:

≤ 130 PSEC

c) Jitter, stability:

OK

d) Prime power:

110 / 220 V

50 - 60 Hz

46

10 VOLTS / DIV.

1.0 NSEC / DIV.

400 KHz.