


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-F*

S.N.: *1078*

Date: *04T 17 82*



- a) Output signal amplitude:
≥ +5 VOLTS
- b) Pulse width:
2 TO 100 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE FIG 1 + 2
- f) Jitter, stability:
OK
- g) Prime power:

1.0 NSEC/DIV
5 VOLTS/DIV
40 KHZ.

+15 ± 1.0 V
250 mA.