


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-IP

S.N.: 1084

Date: JAN 12 83



- a) Output signal amplitude:
0 TO +350V
- b) Pulse width:
0 TO 100 nsec
- c) Rise time:
 ≈ 1.5 nsec
- d) Fall time:
 ≤ 3 nsec
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 320 V/DIV
10 nsec/DIV
2 KHz

+15 \pm 1.0 V
450 mA max.

[Signature]