


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT-N

S.N.: 1088

Date: DEC 17 82



a) Output signal amplitude:

$\approx -30$  VOLTS

b) Pulse width:

$\leq 1$  NSEC

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$  V

150 mA (MAX)

$\approx 32$  VOLTS / DIV.

1.0 NSEC / DIV.

406 KHz