


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT2 -N

S.N.: 1090

Date: DEC 17 82



$\approx 32 \text{ VOLTS/DIV}$
 2.0 NSER/DIV
 400 KHz

a) Output signal amplitude:

$\approx -60 \text{ V}$

b) Pulse width:

$\leq 2 \text{ NSER}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ V}$

150 mA (MAX)