



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-IBM-1*
S.N.: *1099*
Date: *DEC 30 62*



P_{out} *10 VOLTS/DIV*
 1.0 NSEC/DIV
 4.0 MHz



N_{out} *10 VOLTS/DIV*
 1.0 NSEC/DIV
 4.0 MHz

- a) Output signal amplitude:
0 TO ±5V
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24V ± 1.0V
300 mA (MAX)

[Signature]