


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-3-C-D-P*

S.N.: *1100*

Date: *JAN 31 83*



a) Output signal amplitude:

*0 TO +15V*

b) Pulse width:

*0 TO 15 NS*

c) Rise time:

*150*

d) Fall time:

*600*

e) PRF:

*0-25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*110/220V*

*50-60 Hz.*

*10 VOLTS/DIV*

*2 NS/DIV*

*5 MHz*

*[Signature]*