


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-1-PW-N-250

S.N.: 1101

Date: OCT 31 82



- a) Output signal amplitude:  
0 TO -250 V (50  $\Omega$ )
- b) Pulse width:  
100 NSEC TO 5  $\mu$ sec.
- c) Rise time:  
 $\leq$  50 NSEC
- d) Fall time:  
 $\leq$  50 NSEC.
- e) PRF:  
0 TO 5 KHZ
- f) Jitter, stability:  
DIC
- g) Prime power:

50 VOLTS / DIV.  
200 NSEC / DIV.  
1 KHZ.  
 $R_L = 50 \Omega$

120 / 240 V  
50 - 60 HZ.