


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-1-PW-N-250*

S.N.: *1102*

Date: *OCT 31 82*



50 VOLTS/DIV.

200 NSEC/DIV.

1 KHZ.

R_L = 50 Ω

- a) Output signal amplitude:
0 TO -250 V (50 Ω)
- b) Pulse width:
100 NSEC TO 5 μ sec
- c) Rise time:
 \leq 50 NSEC
- d) Fall time:
 \leq 50 NSEC
- e) PRF:
0 TO 5 KHZ.
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz.

[Signature]