


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: 19VH-MPT-N

S.N.: 1105

Date: DEC 17 82



a) Output signal amplitude:

$\approx -30V$

b) Pulse width:

$\leq 1 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0V$

115 mA MAX

$\approx 32 \text{ VOLTS/DIV}$

1.0 nsec/DIV

400 kHz