

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


AVH - AMP T 2 - N

S.N.:

1106

Date:

DEC 17 82



a) Output signal amplitude:

$\geq -60V$

b) Pulse width:

$\leq 2 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0V$

130 mA MAX

[Signature]

$\approx 32 \text{ VOLTS/DIV}$

2.0 nSEC/DIV