


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-P*

S.N.: *1110*

Date: *NOV 29 82*



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0V
200 mA mA

10 V / DIV
1.0 N SEC / DIV
400 KHz