


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-02*

S.N.: *1113*

Date: *JAN 6 82*



- a) Output signal amplitude:  
*0 TO - 40 AMP*
- b) Pulse width:  
*30 NSEC TO 150 NSEC*
- c) Rise time:  
*≈ 5 NSEC*
- d) Fall time:  
*≈ 5 NSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: \*

*50 NSEC/DIV*

*32 VOLTS/DIV*

*4 KHz*

*R<sub>L</sub> = 0.72 Ω SITUATED  
BY 50Ω CABLE  
TO SAMPLING  
SCOPE.*

*120/240 V*

*50 - 60 Hz*

*\* UNIT SHIPPED  
WITH VOLTAGE  
SELECTOR SWITCH  
IN THE 240V  
POSITION*