


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-P

S.N.: 1128

Date: DEC 30 82



- a) Output signal amplitude:
10 TO 50 AMPS
- b) Pulse width:
≤ 25 NSEC
- c) Jitter, stability:
OK
- d) Prime power:

+15 ± 1.0 V.

150 mA (max)

[Handwritten signature]

≈ 20 AMPS / DIV.
10 NSEC / DIV
1 KHZ