


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVP-AV-15-C-P-05*

S.N.: *1133*

Date: *JAN 31 83*



*5 VOLTS / DIV
1.0 NSEC / DIV
100 KHz*



*5 VOLTS / DIV
100 PSEC / DIV
100 KHz*

- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 4 NSEC.
- c) Rise time:
≤ 60 PSEC.
- d) Fall time:
≤ 0.1 NSEC.
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz.*

*UNIT SHIPPED
WITH VOIFAGE
SERVO CARD
IN THE 240V
POSITION.*



[Signature]