


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-HV3-P*

S.N.: *1134*

Date: *NOV 30 82*



a) Output signal amplitude:

*> +60 VOLTS*

b) Pulse width:

*≤ 1.0 NSEC AT 10% RISE POINT*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 V*

*250 mA (MAX)*

*32 VOLTS/DIV  
1.0 NSEC/DIV  
400 KHZ*