


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-T*
S.N.: *1138*
Date: *NOV 29 82*



- a) Output signal amplitude:
0 TO 7.5V
- b) Pulse width:
0 TO 4 NSOR
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V
200 mA (MAX)

10 V / DIV
1.0 NSOR / DIV
400 KHz

[Signature]