


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-D-N-C*

S.N.: *1140*

Date: *JAN 31 83*



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 NSEC/DIV.
4.0 MHz*

*120/240 V,
50-60 Hz.*

*UNIT SKIPPED
WITH VOLTAGE
SELECTOR CARD
IN THE 240V
POSITION.*