


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-2-A-P

S.N.: 1141

Date: MAR 14 83



- a) Output signal amplitude:  
0 TO +5 AMPS
- b) Pulse width:  
0 TO 50 NSEC
- c) Rise time:  
≤ 1 NSEC
- d) Fall time:  
≤ 2 NSEC
- e) PRF:  
0 TO 20 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:

+15 ± 1.0 V.

250 mA (MAX)

5 NSEC/DIV  
≈ 7 AMPS/DIV  
10 KHZ.