


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVPM-2-N

S.N.: 1146

Date: Jan 30 83



- a) Output signal amplitude:
0 TO 15V
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 M/SEC
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0V
500 mA (MAX)

10 VOLTS/DIV
1.0 NSEC/DIV
4.0 M/SEC