


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM-2-N*

S.N.: *1149*

Date: *JAN 30 83*



*10 VOLTS/DIV*

*1.0 nSEC/DIV*

*4.0 MHz*

a) Output signal amplitude:

*0 TO 15V*

b) Pulse width:

*0 TO 2 nSEC*

c) Rise time:

*≤ 100 pSEC*

d) Fall time:

*≤ 135 pSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 V*

*500 mA (MAX)*

*[Signature]*