


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-2-A-N*

S.N.: *1150*

Date: *MAR 14 83*



- a) Output signal amplitude:
0 TO -5 AMPS
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*= 7 AMPS/DIV
5.0 NSEC/DIV
10 KHZ.*

*+15 ± 1.0V
225 mA MAX*